



# EMSL Analytical, Inc.

200 Route 130 North Cinnaminson, NJ 08077  
Phone/Fax: (800) 220-3675 /  
<http://www.EMSL.com> / [IndustrialHygienelab@emsl.com](mailto:IndustrialHygienelab@emsl.com)

EMSL Order ID:      Sample Number  
Customer ID:        Sample Number  
Customer PO:  
Project ID:

**Attn:** MichiganPro.com llc      Phone: 800-470-7938  
Corporate Offices      Fax:  
5632 East Blue Gill Bend      Collected:  
Ellsworth, MI 49729      Received: 6/29/2018  
      Analyzed: 6/29/2018

**Proj:**

### Test Report:Wipe Analysis for Methamphetamine residue by ESI-LC/MS using modified NIOSH 9111

Sample ID	Identification	Area	Sample Weight	Sample Concentration	Reporting Limit
101 281802930-0001	Kitchen Light	100 cm <sup>2</sup>	<0.10 µg	<0.10 µg/100 cm <sup>2</sup>	0.10 µg/100 cm <sup>2</sup>
102 281802930-0002	Duct	100 cm <sup>2</sup>	<0.10 µg	<0.10 µg/100 cm <sup>2</sup>	0.10 µg/100 cm <sup>2</sup>
103 281802930-0003	Living Room Mantel	100 cm <sup>2</sup>	<0.10 µg	<0.10 µg/100 cm <sup>2</sup>	0.10 µg/100 cm <sup>2</sup>
104 281802930-0004	Dining Room Light	100 cm <sup>2</sup>	<0.10 µg	<0.10 µg/100 cm <sup>2</sup>	0.10 µg/100 cm <sup>2</sup>
105 281802930-0005	Family Room Sill	100 cm <sup>2</sup>	<0.10 µg	<0.10 µg/100 cm <sup>2</sup>	0.10 µg/100 cm <sup>2</sup>
Lab Blank		N/A	<0.10 µg	<0.10 µg	N/A

N/A = Not Applicable

#### Analyst(s)

Alicia Shafer

No discernable Field Blanks in this sample set.

Scott Van Etten, CIH, Laboratory Manager

Any questions please contact Scott VanEtten.

Initial report from: 07/06/2018 15:15:08

1. Samples were received in acceptable condition unless otherwise noted.  
2. These results relate only to the samples tested.  
3. Sample results are not blank corrected.

Samples analyzed by EMSL Analytical - Industrial Hygiene Cinnaminson, NJ AIHA-LAP, LLC--IHLAP Accred. Lab 100194